




REV	DESCRIPTION	ECN #	CHG BY	DATE	APP'D
A	ADDED ITEM 22, 2" BIOHAZARD LABEL	PJCT288	JR	07/29/15	CRW
B	UPDATED GAUGE MOUNTING	-	JSR	05/18/16	CRW
C	UPDATED DRAIN	-	JEB	03/18/19	DDH
D	UPDATED DRAIN TO FIT BUCKET	18554	RWK	02/28/20	EZR
E	UPDATED DRAIN MATERIAL	DCR 153	NAS	02/23/21	NAD
F	ADDED LABEL & NOTES	21296	NAS	03/08/21	NAD
G	REMOVED NOTE	PRJ 682	NAS	03/22/21	EZR



ELEMENT SPECS		
PART NUMBER	UL234/2G	
MATERIAL	ULPA	
FLOW RATE	217 CFM	369 m <sup>3</sup> /hr
SURFACE AREA	22.1 ft <sup>2</sup>	2.1 m <sup>2</sup>
I.D.	4.75"	121mm
O.D.	7.75"	197mm
HEIGHT	9.51"	242mm

MODEL #: HV-UL234/2G-301C		MATERIAL: VARIOUS		FINISH: VARIOUS	
MODEL NUMBER: HV-UL234/2G-301C					SHEET: 1 OF 1
DESCRIPTION: HV-UL234/2G-301C, BOTTLE, BALL VALVE, PD GAUGE					
PROJECT:		DRAWN BY: CRW DATE: 5/18/2015		APPROVED BY: M.C. DATE: 5/18/2015	
				REV: G	